


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 240123A/N-104	Page	: 1 of 1
CALIBRATED ON.	: 23 / 01 / 2024	Ref. Std.	: ISO:8655:6
CALIBRATION DUE.	: 22 / 01 / 2025	C.P. No.	: ACS/CP/08
SRF No. & DATE	: 240123A/N Dt. 23.01.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342724000000113F Dt. 27.01.2024	Cal. At	: Site / Lab

CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Ground Floor Shop No 1, Leela Health Care, Kailash Nagar, Majura Gate, Surat, Gujrat 395001
-----------------------	--

ITEM CALIBRATED	: Micro Pipette
Make	: ERBA
Range	: (10 to 100) µl
L. Count	:
ID No.	: RCL/BIO/PIP-04
Location	: Clinical Biochemistry
S. No.	: YE213AS0158493

ENVIRONMENT CONDITION:	Temp. { 25°C ± 4°C } R.H. {30% to 75%}
-------------------------------	--

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	9.57	0.43
100		50	49.43	0.57
		100	99.53	0.47

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.4 µl at coverage factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

