



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - OCTOBER 2023



PC-1024

Lab Name	Govt. PHC, Asanur	Lab No	13138
Constituent Group	Chemistry III	Date of Result Entered :	20/10/2023
PT Item	Lyophilized human serum based	Date of Report Published :	04/11/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1905	118.63	6.44	7.63	96 mg/dL	-2.96	0.35
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1912	7.12	5.02	0.36	6.8 mg/dL	-0.90	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1786	111.59	4.28	4.78	109 mg/dL	-0.54	0.23

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :
Govt. PHC, Asanur
Asanur
TAMILNADU

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Pamela Christudoss
மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலையம்,
ஆசனூர்
உளுந்துவாடி, புலகுடி வட்டம்,