



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - NOVEMBER 2023



Lab Name	Govt. PHC, Asanur	Lab No	13138
Constituent Group	Chemistry III	Date of Result Entered :	20/11/2023
PT Item	Lyophilized human serum based	Date of Report Published :	05/12/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1894	218.58	5.36	11.73	229 mg/dL	0.89	0.54
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1871	4.36	6.15	0.27	4.5 mg/dL	0.52	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1778	106.36	5.29	5.62	116 mg/dL	1.71	0.27

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

Govt. PHC, Asanur

Asanur

TAMILNADU

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christlan Medical College, Vellore

Homogeneity and Stability of the sample Is passed.

Data In CMC EQAS reports Is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

[Signature]
மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலையம்,
ஆசனூர்
உளுந்துார்-பட்டை வட்டம்.