



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2023



PC-1024

Lab Name Govt. PHC, Asanur Lab No 13138
 Constituent Group Chemistry III Date of Result Entered : 20/12/2023
 PT Item Lyophilized human serum based Date of Report Published : 02/01/2024

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1885	99.99	6.05	6.05	89 mg/dL	-1.82	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1844	1.74	6.08	0.11	1.7 mg/dL	-0.38	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	107.76	4.90	5.28	101 mg/dL	-1.28	0.25

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :
 Govt. PHC, Asanur
 Asanur
 TAMILNADU

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

[Handwritten Signature]
 மருத்துவ அறிவியல்
 அரசு ஆரம்ப சுகாதார நிலையம்,
 ஆசனூர்
 உருத்தாரப்பேட்டை வட்டம்.