



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JANUARY 2024



Lab Name	Govt. PHC, Asanur	Lab No	13138
Constituent Group	Chemistry III	Date of Result Entered :	20/01/2024
PT Item	Lyophilized human serum based	Date of Report Published :	06/02/2024

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1882	145.08	6.30	9.14	142 mg/dL	-0.34	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1872	1.64	7.16	0.12	1.6 mg/dL	-0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1799	107.07	5.63	6.03	101 mg/dL	-1.01	0.28

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :  
Govt. PHC, Asanur  
Asanur  
TAMILNADU

Coordinator Contact Details:  
Email:cllnqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*[Handwritten Signature]*  
மருத்துவ கல்வி நிறுவனம்,  
சிரசு ஆரம்ப கல்வி நிறுவனம்,  
சுசுவா  
உளுந்துப்பட்டை வட்டம்,