

CALIBRATION CERTIFICATE

Page No:1 of 1

CERTIFICATE NO: SBS/CL/23/13622		MEDICAL DEVICES		
Issue Date	07-10-2023			
SRF No & Date	SRF/23/00783-0001 & 07-10-2023			
Receipt Date	07-10-2023			
Calibration Date	07-10-2023			
Calibration Due	06-10-2024			
Customer Name & Address				
GOVERNMENT PRIMARY HEALTH CENTRE, ASANUR,KALLAKURICHI DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	SEMI AUTO ANALYZER	Make & Model	ROBONIK & PRIETEST TOUCH	
Range	MULTI	Sr. No	ATCD2850321RBK	
Resolution	MULTI	Identification No	NA	
DUC Condition	SATISFACTORY	Location	LABORATORY	
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature:25.6 ° C	Relative Humidity	52% RH	
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance >20MΩ	Measured values in MΩ 88	Uncertainty in % (±) 13.92
2	Earth Leakage <5000μAfor B,BF,CF	Measured values in μA 175	Uncertainty in % (±) 7.5
3	Enclosure Leakage <500μAfor B,BF,CF	Measured values in μA 235	Uncertainty in % (±) 7.2

REMARKS

- 1 This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2 The user should determine the suitability of the instrument for its intended use.
- 3 The recalibration interval should be determined on the user requirement.
- 4 The results stated in this certificate relate only to the item calibrated.
- 5 The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00
- 6 Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

KSTy
(Calibration Engineer)
K SATHYAMOORTHY



Authorised Signatory

C.SL
Technical Manager
C.SHANMUGARAJ

Chief Executive