



Metropolis Healthcare Ltd., EQAS

PRIMUS PATHOLOGY & DIAGNOSTICS (1015)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 04

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Nihon Kohden Celltac Alpha

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	1.50	1.54	0.37	0.16
@ Total Leucocytes (WBC) count	/c.mm	4800.00	4530.00	0.23	-0.24
@ Haemoglobin	g/dL	4.20	4.39	-0.93	-1.14
@ MCV (Mean Corpuscular Volume)	fL	131.33	128.57	0.13	0.13
@ MCH (Mean Corpuscular Hb)	pg	28.00	28.47	-0.54	-1.02
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	21.32	22.24	-0.59	-0.84
@ Platelet Count	1000/ µL	140.00	141.67	-0.34	-0.24
@ Hematocrit	%	19.70	19.80	0.17	0.11
@ RDW CV%	%	38.30	32.60	0.81	0.87

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score ≥ 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

PRIMUS PATHOLOGY & DIAGNOSTICS (1015)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 03

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Nihon Kohden Celltac Alpha

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	1.04	1.09	-0.05	0.16
@ Total Leucocytes (WBC) count	/c.mm	2400.00	2548.00	-0.70	-0.24
@ Haemoglobin	g/dL	2.50	2.90	-1.34	-1.14
@ MCV (Mean Corpuscular Volume)	fL	130.77	127.61	0.13	0.13
@ MCH (Mean Corpuscular Hb)	pg	24.04	26.71	-1.49	-1.02
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	18.38	21.10	-1.09	-0.84
@ Platelet Count	1000/ µL	74.00	77.93	-0.13	-0.24
@ Hematocrit	%	13.60	13.87	0.05	0.11
@ RDW CV%	%	36.30	32.35	0.93	0.87

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

PRIMUS PATHOLOGY & DIAGNOSTICS (1015)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 01

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____


Reviewed by: _____

Dated: _____

Instrument : Standard F 200

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	6.70	6.70	-1.50	-1.39

Legend @ : Acceptable	Total Parameters	1
! : 2.0 < z score < 3.0 - Warning	Not Evaluated Parameters	0
X : z score ≥3 - Unacceptable	Evaluated Parameters	1
# : Not Evaluated	Outlier Parameters (X)	0
⌚ : Delayed Result Entry	EQAS Score HbA1c	100.00 %
* : Not considered for evaluation.		



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroil Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

PRIMUS PATHOLOGY & DIAGNOSTICS (1015)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 02

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____


Corrective Action: _____

Reviewed by: _____ **Dated:** _____

Instrument : Standard F 200

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	5.50	5.48	-1.27	-1.39

Legend @ : Acceptable	Total Parameters	1
! : 2.0 < z score < 3.0 - Warning	Not Evaluated Parameters	0
X : z score ≥3 - Unacceptable	Evaluated Parameters	1
# : Not Evaluated	Outlier Parameters (X)	0
⌚ : Delayed Result Entry	EQAS Score HbA1c	100.00 %
* : Not considered for evaluation.		



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroil Road, Kurla (W),
Mumbai - 400070