



**Department of Clinical Biochemistry
Christian Medical College, Vellore**



CMC EXTERNAL QUALITY ASSURANCE SCHEME

(Under the aegis of ACBI)

PC-1024

Certificate of Participation

This is to certify that

REGIONAL AYURVEDA RESEARCH INSTITUTE

VIJAYAWADA

Lab No.18753..... has participated

*in the HbA1c Program of the External Quality Assurance Scheme
during the year 2023.*

Pamela Christudoss

Dr. Pamela Christudoss, M.Sc., Ph.D.,

EQAS Coordinator

Professor of Clinical Biochemistry



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



Lab Name **REGIONAL AYURVEDA RESEARCH INSTITUTE** Lab No **18753**
Constituent Group **HbA1c** Date of Result Entered : **14/07/2023**
PT item **Lyophilized human whole blood based** Date of Report Published : **07/08/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	297	7.34	10.92	0.80	8 %	0.82	0.09

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

**REGIONAL AYURVEDA RESEARCH INSTITUTE
NEW RAJEEV NAGAR, PAYAKAPURAM
VIJAYAWADA
ANDRA PRADESH 520015**

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name **REGIONAL AYURVEDA RESEARCH INSTITUTE** Lab No **18753**
Constituent Group **HbA1c** Date of Result Entered : **16/08/2023**
PT item **Lyophilized human whole blood based** Date of Report Published : **01/09/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	304	6.41	9.14	0.59	6.9 %	0.84	0.07

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

**REGIONAL AYURVEDA RESEARCH INSTITUTE
NEW RAJEEV NAGAR, PAYAKAPURAM
VIJAYAWADA
ANDRA PRADESH 520015**

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name **REGIONAL AYURVEDA RESEARCH INSTITUTE** Lab No **18753**
Constituent Group **HbA1c** Date of Result Entered : **08/09/2023**
PT item **Lyophilized human whole blood based** Date of Report Published : **05/10/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	280	7.80	10.95	0.85	8.5 %	0.82	0.10

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

**REGIONAL AYURVEDA RESEARCH INSTITUTE
NEW RAJEEV NAGAR, PAYAKAPURAM
VIJAYAWADA
ANDRA PRADESH 520015**

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - OCTOBER 2023



PC-1024

Lab Name **REGIONAL AYURVEDA RESEARCH INSTITUTE** Lab No **18753**
Constituent Group **HbA1c** Date of Result Entered : **12/10/2023**
PT item **Lyophilized human whole blood based** Date of Report Published : **04/11/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	265	6.57	9.81	0.64	7.4 %	1.29	0.08

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

**REGIONAL AYURVEDA RESEARCH INSTITUTE
NEW RAJEEV NAGAR, PAYAKAPURAM
VIJAYAWADA
ANDRA PRADESH 520015**

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - NOVEMBER 2023



PC-1024

Lab Name **REGIONAL AYURVEDA RESEARCH INSTITUTE** Lab No **18753**
Constituent Group **HbA1c** Date of Result Entered : **16/11/2023**
PT item **Lyophilized human whole blood based** Date of Report Published : **05/12/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	251	7.64	11.81	0.90	7.7 %	0.07	0.11

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

**REGIONAL AYURVEDA RESEARCH INSTITUTE
NEW RAJEEV NAGAR, PAYAKAPURAM
VIJAYAWADA
ANDRA PRADESH 520015**

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2023



PC-1024

Lab Name **REGIONAL AYURVEDA RESEARCH INSTITUTE** Lab No **18753**
Constituent Group **HbA1c** Date of Result Entered : **18/12/2023**
PT item **Lyophilized human whole blood based** Date of Report Published : **02/01/2024**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	302	6.10	10.60	0.65	6.7 %	0.93	0.07

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

**REGIONAL AYURVEDA RESEARCH INSTITUTE
NEW RAJEEV NAGAR, PAYAKAPURAM
VIJAYAWADA
ANDRA PRADESH 520015**

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JANUARY 2024



PC-1024

Lab Name **REGIONAL AYURVEDA RESEARCH INSTITUTE** Lab No **18753**
Constituent Group **HbA1c** Date of Result Entered : **18/01/2024**
PT item **Lyophilized human whole blood based** Date of Report Published : **06/02/2024**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	294	6.18	12.82	0.79	7 %	1.03	0.09

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

**REGIONAL AYURVEDA RESEARCH INSTITUTE
NEW RAJEEV NAGAR, PAYAKAPURAM
VIJAYAWADA
ANDRA PRADESH 520015**

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****