



APPENDIX E

THERMOMETER CALIBRATION SHEET

Calibration of Thermometer

Name and ID of equipment: Refrigerator -01 (ICTC - IGPC)

ID of Thermometer to be calibrated: ICTC - T-1

Temperature range of thermometer: -10°C to +11°C

Name and ID of Calibrator: Cal T-2

Temperature range of Calibrator: -10°C to +11°C

Thermometer calibration done on: 11/01/2024

Thermometer calibration due on: 10/01/2025

Thermometer calibration done at: HIV Lab, MAMC

Time of reading	Temperature of Working Thermometer	Temperature of Calibrator	Difference in readings
After 1 hour	4°C	4°C	0
After 2 hour	4°C	4°C	0
After 3 hour	4°C	4°C	0

Correction Factor: 0

Comments: Nil

Calibrated By: Ramesh Kumar

Bhush
11/01/2024

Technical Officer

HIV LABORATORY
Department of Microbiology
Maulana Azad Medical College
New Delhi-110002

Document Type: SOP for operation and monitoring of Refrigerator	Document No: MAMC/HIV/SOP/07	
Issue No.:02	Issue Date: 10.09.2016	Page No.: 1 of 1