



APPENDIX C

Centrifuge Calibration Sheet ICTC IGPC

Name and Serial No. of Equipment: Centrifuge-01(S.No.-202012235-12) ElectraBL

Name and ID of Calibrator: Tachometer (Conitronics) (HIV Lab MAMC) (12 T)

Centrifuge Calibration done on: 18/01/2024

Centrifuge Calibration due On: 17/01/2025

S. No.	Set Limit	Acceptable range	Tachometer reading
1	1500	1530 to 1470	1510
			1490
			1480
2	3000	3060 to 2940	3050
			3010
			3030
3	4500	4590 to 4410	4510
			4480
			4470

Verification of rotation speed:

If the difference between the test centrifuge and the certified tachometer is $\pm 2\%$ of the procedure speed, then the test centrifuge rotation calibration is verified as acceptable.

Result: Centrifuge Performing within the range

Comments: Nil

Calibrated By: Priya Shukla (T.O. HIV Lab, MAMC)

Shukla
18/01/2024

Technical Officer

Document Type: SOP for Operation and Monitoring Centrifuge	Document No: MAMC/HIV/SOP/09 of Microbiology	HIV LABORATORY Maulana Azad Medical College New Delhi-110002
Issue No.:02	Issue Date: 10.09.2015	