



Neu-QAP

(an EXTERNAL QUALITY ASSURANCE PROGRAMME)

PARTICIPANT FINAL ASSESSMENT REPORT

PT SCHEME: SEROLOGY



PC-1047

Cycle No	C5
Ref.No.	NEUQAP610
Sample ID	NEUQAP SERO SERUM/2023/10/C5/S3

Report Date : 28/12/2023

Sample : October 2023

Peer group

Parameters	Lab Report	No of Participants	Consensus Report	%	Remarks
HBsAg (Card Method)	Reactive	15	Non Reactive - 1 Reactive - 14	93	Within Consensus
HCV (Card method)	Equivocal	16	Non Reactive - 4 Reactive - 11 Equivocal - 1		Out of Consensus
HIV-1 & 2 (Card method)	Reactive	14	Non Reactive - 1 Reactive - 13	93	Within Consensus
RPR (Others)	Non Reactive	4	Non Reactive - 3 Reactive - 1	75	Within Consensus
Widal (Slide agglutination)	Negative	25	Negative - 25	100	Within Consensus

Authorised Signatory

Dr. Sujay Prasad

Technical Manager and Program coordinator

Technical Advisor

Dr Divya

Page 2 of 4

This report is for use only by the intended participant.



An external quality assurance program from
NEUBERG ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.

Anand Tower, 54, Bowring Hospital Road,
Shivajinagar, Bangalore - 560001
+91-9902277355 · qap@naalm.com · www.naalm.com
CIN: U85300KA2018PTC115147