

Date: 10/09/ 2023

Certificate of Calibration

Name & Address of Customer: Shiv Pathology Lab

G3-G4, Lajpat Nagar, (Near Shyam Park Metro Station) Sahibabad

City: Sahibabad .State- Uttar Pradesh, PIN: 201005

Phone: 9953048485, E-Mail:shiva.plab@gmail.com

Name / Model of Instrument: Fully auto Biochemistry Analyzer

Type: Astira C-MS (Q Line biotech PVT LTD).

Serial No: 220T180CS0031S

Calibration Date: 10/09/ 2023

Valid Date: 09/09/2024

Next Calibration Due: 10/09/ 2024

This is to certify that above said instrument has been calibrated for aspiration/dispensing, counting time & LED Current according to the procedures provided by Q line biotech PVT LTD.

These instruments conform to, CE-IVD & EU directives of use.

Calibration carried out on site by



Signature

Name of Engineer: Rajiv Ranjan

Instrument Service Division

Q line biotech Pvt. Ltd

Q-LINE BIOTECH PRIVATE LIMITED (Formerly known as "POCT Services Pvt. Ltd.")

Corp. Office: 20A (Basement), Main Shivaji Marg, Najafgarh Road, Near Moti Nagar Police Station, New Delhi-110015, INDIA. Tel.: 011-45577407

Regd. Office: 298-281, Transport Nagar, Kanpur Road, Adjacent Transport Nagar Metro Station, Lucknow-226012 (U.P.), INDIA. Tel.: 0522-2433023

E-mail: info@qlinebiotech.com | Website: www.qlinebiotech.com | CIN No.: U74120UP2010PTC042528

