

## CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12474	MEDICAL DEVICES	Page No 1 of 1
Issue Date	30-09-2023	
SRF No & Date	SRF/23/00545-0001 & 29-09-2023	
Receipt Date	29-09-2023	
Calibration Date	29-09-2023	
Calibration Due	28-09-2024	

**Customer Name & Address**  
 GOVERNMENT PRIMARY HEALTH CENTRE,  
 PAPANKULAM-642204, TIRUPPUR DISTRICT

**Details of Device Under Calibration (DUC)**

Description	SEMI AUTO ANALYZER	Make & Model	NA & PRITEST TOUCH
Range	MULTI	Sr No	AT4020211RBK
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

**Environmental Conditions & Calibration Procedure Details**

Environmental Details	Temperature 25.6°C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

**Reference Standards Details**

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-05-2024


**ELECTRICAL SAFETY**


**RESULTS**

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 90	Uncertainty in % (±) 13.92
2	Earth Leakage <6000µAfor B,BF,CI	Measured values in µA 195	Uncertainty in % (±) 7.4
3	Enclosure Leakage <600µAfor B,BF,CI	Measured values in µA 232	Uncertainty in % (±) 7.2

**REMARKS**

- 1 This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory
- 2 The user should determine the suitability of the instrument for its intended use.
- 3 The recalibration interval should be determined on the user requirement
- 4 The results stated in this certificate relate only to the item calibrated
- 5 The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00
- 6 Equipment used for Calibration were calibrated & traceable to National & International Standards

Calibrated By  
  
 (Calibration Engineer)  
 M RAGUL

  
 Technical Manager  
 C. SHANMUGARAJ

Chief Executive

