



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JANUARY 2024



Lab Name **DELVIN PATHOLOGY LABORATORY** Lab No **18304**  
Constituent Group **Chemistry II** Date of Result Entered : **13/01/2024**  
PT item **Lyophilized human serum based** Date of Report Published : **06/02/2024**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	969	152.52	9.94	15.15	159 mg/dL	0.43	0.97
2	UREA II	Urease UV / GLDH II	Transasia / Erba	747	46.49	10.32	4.80	53 mg/dL	1.36	0.35
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	608	1.69	11.03	0.19	1.88 mg/dL	1.02	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	823	3.31	11.52	0.38	3.52 mg/dL	0.55	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	731	5.13	8.46	0.43	5.4 g/dL	0.62	0.03
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	807	3.12	9.42	0.29	3.1 g/dL	-0.07	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	850	7.84	12.11	0.95	7.8 mg/dL	-0.04	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	849	100.98	11.05	11.16	100 mg/dL	-0.09	0.77
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	804	132.15	12.98	17.16	130 mg/dL	-0.13	1.21

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

**DELVIN PATHOLOGY LABORATORY**  
Shop No FF01 Second Floor, Parsvnath Plaza, G Block, Pocket G, Sector 27  
NOIDA  
UTTAR PRADESH 201301

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*