



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JANUARY 2024



Lab Name **MICRO HEALTH LABORATORIES PRIVATE LIMITED**

Lab No **19585**

Constituent Group **Chemistry I**

Date of Result Entered : **15/01/2024**

PT item **Lyophilized human serum based**

Date of Report Published : **06/02/2024**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE	Hexokinase	Beckman AU480/680/5800/DXC700AU	142	123.19	4.67	5.76	133 mg/dL	1.70	0.97
2	UREA	Urease UV / GLDH	Beckman AU480/680/5800/DXC700AU	161	36.78	4.27	1.57	37 mg/dL	0.14	0.25
3	CREATININE	Enzymatic Colorimetric	Beckman AU480/680/5800/DXC700AU	36	1.54	7.66	0.12	1.6 mg/dL	0.51	0.04
4	T.BILIRUBIN	Others (DPD, Vanadate Oxidation)	Any Analyser (Automation / Semi Automation)	225	2.21	14.32	0.32	2.5 mg/dL	0.91	0.04
5	T-PROTEIN	Biuret - Colorimetric	Beckman AU480/680/5800/DXC700AU	164	4.79	3.53	0.17	4.6 g/dL	-1.12	0.03
6	ALBUMIN	BCG - colorimetric	Beckman AU480/680/5800/DXC700AU	163	3.08	4.22	0.13	3.1 g/dL	0.15	0.02
7	CALCIUM	Arsenazo III	Beckman AU480/680/5800/DXC700AU	159	8.82	3.71	0.33	8.9 mg/dL	0.24	0.05
8	URIC ACID	Enzymatic / Uricase Colorimetric	Beckman AU480/680/5800/DXC700AU	161	4.60	3.72	0.17	4.7 mg/dL	0.58	0.03
9	CHOLESTEROL	CHOD-PAP	Beckman AU480/680/5800/DXC700AU	162	107.68	4.72	5.08	117 mg/dL	1.83	0.80
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Beckman AU480/680/5800/DXC700AU	160	143.47	5.15	7.39	144 mg/dL	0.07	1.17
11	HDL	Direct method / Enzymatic colorimetric	Beckman AU480/680/5800/DXC700AU	147	24.51	8.05	1.97	21 mg/dL	-1.78	0.33
12	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1155	130.94	3.25	4.26	133 mmol/L	0.48	0.25
13	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1188	3.20	5.97	0.19	3.4 mmol/L	1.05	0.01
14	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	969	95.29	4.58	4.37	97 mmol/L	0.39	0.28
15	AST	UV kinetic(with & without PLP (P-5-P))	Beckman AU480/680/5800/DXC700AU	157	113.27	5.84	6.62	129 U/L	2.38	1.06
16	ALT	UV kinetic(with & without PLP (P-5-P))	Beckman AU480/680/5800/DXC700AU	156	43.63	9.47	4.13	47 U/L	0.82	0.66
17	ALP	PNP AMP kinetic	Beckman AU480/680/5800/DXC700AU	154	202.30	8.99	18.19	209 U/L	0.37	2.93

u* - Method of Uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

**MICRO HEALTH LABORATORIES PRIVATE LIMITED
NEAR GOVT MEDICAL COLLEGE, THAZHEANGADI ROAD, MANANTHAVADY
WAYANAD
KERALA670645**

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components

***** End of Report *****