



N.K Engineering & Research Center

Plot No. 142, Sneh Bharti Place, Mangal Chowk, Khemnichak, Ram Krishna Nagar, Patna, Bihar-800027
Mob: 8766317650, 7838663310

Website : www.nkerc.com, E-mail: info@nkerc.com, info.nkerc@gmail.com

Calibration Certificate

Page 1 of 1

Format No.-QS-22F

Certi. No.	NKERC/23-24/1906	Calibration Date	28.02.2024	S.R.F.No & Date	95/26/02/2024
Issue Date	29.02.2024	Sug.Cali. Due Date	27.02.2025	Receipt Date	28.02.2024

Name & Address of Customer:-

M/s:- Biochemistry and Pathology Laboratory.
Regional Research Institute of Unani Medicine,
Guzri, Patna City, Bihar- 800008.

Identification of UUC (Unit Under Calibration)

Instrument Name	Blood Mixer	Visual Condition	Satisfactory
Serial No./ID	325B/17EQ04	Accuracy	-
Make/Model	Transasia	Calibration Performed	Site
Range	20 to 30 RPM	Location	Lab
Least count	-	Reference Standard	-

Calibration Procedure/Method : CI(S)-01

Environmental Condition(s) maintained during calibration :-

Temperature : (23±5)°C Relative Humidity (50±20)%RH

I. DISCIPLINE : Mechanical (Dimension)

Standard & Major Equipment(s) Used:

Sr. No.	Standard Equipments	Range/ ID	Calibration Agency	Certificate No.	Calibration Date	Cal Due date
1	Digital Tachometer	NKERC/DTM-01	Accurate Measurements	AM23CC06239	18.09.2023	18.09.2024

Calibration Results

Sr. No.	UUC Reading (RPM)	Standard Measured Value	Error (RPM)
1	20	20.0	0.0
2	25	25.2	-0.2
3	30	30.1	-0.1

***** END OF REPORT *****

NOTE:

1. The result are valid only for this Calibrated item.
2. The calibration report shall not be reproduced without the written approval of calibration laboratory.
3. Result reported are valid at the time of & under the stated conditions of measurement
4. Laboratory standards are traceable to National/International Standards.
5. Calibration Certificates issued for Weight & Measure parameters i.e. Mass, Balance, Volumetric Equipment, Measuring Scales / Tapes etc. are for scientific purpose only and should not be used for Trade / Commercial use.

CALIBRATED BY


(CALIBRATION ENGINEER)



APPROVED BY


NIKHIL KUMAR JHA
(Q.M/T.M.)