



CERTIFICATE OF CALIBRATION

Certificate No: CC-24-25-S-067/02		Page. No: 1 of 1
Customer Name & Address	ULR No.	CC272524000001211F
M/s. Micro Health Laboratories	SRF No.	VCI-24-25-S-067
Near Kannankandy E-Store,	Date of Receipt	27.02.2024
Mananthavady, Wayanad,	Date of Calibration	27.02.2024
Kerala - 670 645.	Due Date of Calibration	26.02.2025
	Date of Issue	28.02.2024

Details of Office Office	1 Cambration			
Description	Centrifuge	Make	Remi	
Range	0 to 5250 rpm	Model	R8C	
Resolution	NA	Condition of UUC	Good	
Serial Number	ZHBN-05733	Instrument Location		
ID Number	MHL/C/06	Cal. At	On Site	

Environmental Condition		Calibration Method Used		
Temperature (°C)	23±5	National / International Standard	SANAS TR 45-02	
Humidity (%RH)	40 - 70	Cal Procedure No	VCI-L03-WI-SPE-002	

Standard Used

Sl. No.	Description	ID.No.	Certificate No.	Make/Model	Traceability	Valid till
1	Digital Tachometer	VCI-M-TM-068	2023-24/EQN/2307136-01	Line Seiki/TM-5000	Equation CC-2760	24.07.2024

Results of Calibration				
Sl. No.	UUC Reading (rpm)	Standard Reading (rpm)	Error (rpm)	Measurement Uncertainty (±) (%)
01	1000.0	1000.1	-0.1	0.13
02	2000.0	2000.2	-0.2	0.70
03	3000.0	3000.3	-0.3	1.94
04	4000.0	4000.3	-0.3	1.45
05	5250.0	5250.4	-0.4	1.11

- 1. UUC is Defined as the Unit Under Calibration.
- 2. This Calibration Certificate Results relates only to the above UUC & This report shall not be reproduced except in full/part without approval of the Venus Calibration and Instruments.
- 3. The reported results are valid at the time of under stated condition of measurement.
- 4. Calibration of the UUC are traceable to National / International Standards.
- 5. The recalibration interval shall be decided by the user.
- 6. Nabl-133 guidelines are adopted for use of NABL symbol.
- 7. The Measurement uncertainty is expressed at 95.45% confidence level with coverage factor k =2.

Checked & Issued by:

P.Renuka (Engineer - Calibration)

*******END OF THE CERTIFICATE******





(Technical / Quality Manager)