



CC-2725



CERTIFICATE OF CALIBRATION

Certificate No: CC-24-25-S-067/01

Page No : 1 of 1

Customer Name & Address M/s. Micro Health Laboratories Near Kannankandy E-Store, Mananthavady, Wayanad, Kerala - 670 645.	ULR No.	CC272524000001210F
	SRF No.	VCI-24-25-S-067
	Date of Receipt	27.02.2024
	Date of Calibration	27.02.2024
	Due Date of Calibration	26.02.2025
	Date of Issue	28.02.2024

Details of Unit Under Calibration

Description	Refrigerator	Make	TATA
Range	2 to 8°C	Model	Voltas
Resolution	NA	Condition of UUC	Good
Serial Number	KO11007	Instrument Location	--
ID Number	MHL/C/02	Cal. At	On Site

Environmental Condition		Calibration Method Used	
Temperature (°C)	30±5	National / International Standard	DKD-R-5-7
Humidity (%RH)	30 - 75	Cal Procedure No	VCI-L03-WI-THER-005

Standard Used

Sl. No.	Description	ID.No.	Certificate No.	Make/Model	Traceability	Valid till
1	RTD Sensor with Indicator	VCI-T-RTD-005& VCI-ET-DMM-001	CC-24-25-L-005/01	Tempens/Keithley/2110	Venus Calibration CC-2725	05.01.2025

Result of Calibration			
S. No	UUC Setting (°C)	Standard Reading in (°C)	Measurement Uncertainty ± (°C)
1	2 to 8	3.56	0.20

Remarks:

- UUC is Defined as the Unit Under Calibration.
- This Calibration Certificate Results relates only to the above UUC & This report shall not be reproduced except in full/ part without approval of the Venus Calibration and Instruments.
- The reported results are valid at the time of under stated condition of measurement.
- Calibration of the UUC are traceable to National / International Standards.
- The recalibration interval shall be decided by the user.
- Nabl-133 guidelines are adopted for use of NABL symbol.
- The Measurement uncertainty is expressed at 95.45% confidence level with coverage factor k=2.
- Temperature Chart : ITS-90

Checked & Issued by:

P.Renuka
P.Renuka

(Engineer - Calibration)



Authorised by:

C.Dhinakaran
C.Dhinakaran

(Technical / Quality Manager)

*****END OF THE CERTIFICATE*****

No.16, 6th Street, Annai Sathya Nagar, Poothapedu Ramapuram, Chennai - 89.

Ph: +91 44 2476 0533 /0633, Mob: 90031 20649, 9789893643, 8807402558

e: venuscallab@gmail.com. info@venuscal.com | www.venuscal.com