



Metropolis Healthcare Ltd., EQAS

HEALTH KEY PATH LABS PVT. LTD (1130)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230412 and Sample No 04

Report Date : 13/02/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Mindray BC5130

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.21	2.25	-0.07	0.03
@ Total Leucocytes (WBC) count	/c.mm	8390.00	8390.00	0.80	1.19
@ Haemoglobin	g/dL	7.10	7.20	0.25	0.27
@ MCV (Mean Corpuscular Volume)	fL	162.90	158.35	1.42	1.39
@ MCH (Mean Corpuscular Hb)	pg	32.10	31.95	0.27	0.25
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	19.70	20.20	-1.40	-1.41
@ Platelet Count	1000/ μ L	229.00	243.50	1.16	1.97
@ Hematocrit	%	36.00	35.55	1.72	1.77
@ RDW CV%	%	30.10	--	1.67	1.86

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score \geq 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070