



Metropolis Healthcare Ltd., EQAS

PRIMUS PATHOLOGY & DIAGNOSTICS (1015)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 01

Report Date : 09/02/2024

Analyte	Instrument	Result Value	Standard Unit
⌚ # C-Reactive Protein Qualitative	Robonik Prietest Touch	Abnormal	mg/dl
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Manual


Analyte	Standard Unit	Result Value	Accepted Value
⌚@ HBsAg	IU/mL	Non-reactive	Non-reactive
⌚@ HCV	IU/mL	Non-reactive	Non-reactive
⌚@ Rheumatoid Factor Qualitative	IU/mL	Normal	Normal
⌚@ Dengue NS 1	RU/ml	Negative	Negative
⌚@ HIV Ab	S/CO	Non-reactive	Non-reactive
⌚@ Typhidot IgM	-	Negative	Negative
⌚@ Widal test - Typhi O antigen (TO)	-	No Agglutination	1:80,NoA
⌚@ Widal test - Typhi H antigen (TH)	-	No Agglutination	1:80,NoA
⌚@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	1:80,NoA
⌚@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	1:80,NoA

Instrument : Robonik Prietest Touch

Analyte	Standard Unit	Result Value	Accepted Value
⌚# C-Reactive Protein Qualitative	mg/L	Abnormal	

Legend @ : Acceptable
 ⌚@ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	11
Not Evaluated Parameters	1
Evaluated Parameters	10
Outlier Parameters (X)	0
EQAS Score Immunoassay Extended	100.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirool Road, Kurla (W),
 Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

PRIMUS PATHOLOGY & DIAGNOSTICS (1015)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 02

Report Date : 09/02/2024

Analyte	Instrument	Result Value	Standard Unit
⌚ X HIV Ab	Manual	HIV-1/HIV-2 Ab Reactive	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____


Dated: _____

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
⌚@ HBsAg	IU/mL	Reactive	Reactive
⌚@ HCV	IU/mL	Non-reactive	Non-reactive
⌚@ Rheumatoid Factor Qualitative	IU/mL	Normal	Normal
⌚@ Dengue NS 1	RU/ml	Negative	Negative
⌚ X HIV Ab	S/CO	HIV-1/HIV-2 Ab Reactive	HIV-1 Ab Reactive
⌚@ Typhidot IgM	-	Negative	Negative
⌚@ Widal test - Typhi O antigen (TO)	-	No Agglutination	1:80,NoA
⌚@ Widal test - Typhi H antigen (TH)	-	No Agglutination	1:80,NoA
⌚@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	1:80,NoA
⌚@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	1:80,NoA

Instrument : Robonik Prietest Touch

Analyte	Standard Unit	Result Value	Accepted Value
⌚@ C-Reactive Protein Qualitative	mg/L	Abnormal	Abnormal
Legend @ : Acceptable		Total Parameters	11
⌚@ : Acceptable		Not Evaluated Parameters	0
X : Unacceptable		Evaluated Parameters	11
# : Not Evaluated		Outlier Parameters (X)	1
⌚ : Delayed Result Entry		EQAS Score Immunoassay Extended	90.91 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirool Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

PRIMUS PATHOLOGY & DIAGNOSTICS (1015)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 03

Report Date : 09/02/2024

Analyte	Instrument	Result Value	Standard Unit
⌚ # C-Reactive Protein Qualitative	Robonik Prietest Touch	Normal	mg/dl
⌚ X HIV Ab	Manual	HIV-1/HIV-2 Ab Reactive	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Manual


Analyte	Standard Unit	Result Value	Accepted Value
⌚@ HBsAg	IU/mL	Reactive	Reactive
⌚@ HCV	IU/mL	Reactive	Reactive
⌚@ Rheumatoid Factor Qualitative	IU/mL	Normal	Normal
⌚ X HIV Ab	S/CO	HIV-1/HIV-2 Ab Reactive	HIV-1 Ab Reactive

Instrument : Robonik Prietest Touch

Analyte	Standard Unit	Result Value	Accepted Value
⌚ # C-Reactive Protein Qualitative	mg/L	Normal	

Legend @ : Acceptable
 ⌚@ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	5
Not Evaluated Parameters	1
Evaluated Parameters	4
Outlier Parameters (X)	1
EQAS Score Immunoassay Extended	75.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070