

Instrument : Roche Cobas Pure

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Urea	mg/dl	19.80	20.02	-0.04	0.02
@ Creatinine	mg/dl	0.73	0.73	-0.02	0.56
@ Bilirubin Total	mg/dl	0.18	0.18	0.11	0.05
@ Total Protein	g/dL	5.60	5.53	0.03	0.12
@ Albumin	g/dL	3.33	3.35	0.11	0.22
@ Uric Acid	mg/dl	3.71	3.66	-0.05	-0.04
@ Calcium	mg/dl	7.21	7.15	0.12	0.08
@ Phosphorous Inorganic	mg/dl	3.91	3.75	0.73	0.38
@ Cholesterol	mg/dl	129.00	128.57	0.43	0.44
@ Triglycerides	mg/dl	127.00	131.71	-0.28	0.09
@ HDL	mg/dl	26.50	26.64	-1.36	-1.27
@ LDL	mg/dl	68.10	68.13	-0.65	-0.04
@ ALT	U/L	19.00	17.49	0.47	0.47
@ AST	U/L	34.80	33.59	0.81	0.31
@ ALP	U/L	45.60	46.21	0.04	0.26
@ GGT	U/L	24.80	24.42	0.06	0.31
@ Amylase	U/L	59.20	59.35	0.40	0.20
@ Chloride	mmol/L	75.50	74.70	-0.42	-0.25
@ Lipase	U/L	32.90	30.94	0.05	0.33
@ Sodium	mmol/L	109.30	106.77	0.20	0.33
@ Potassium	mmol/L	3.66	3.59	0.31	0.41

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score ≥3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	21
Not Evaluated Parameters	0
Evaluated Parameters	21
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070

Instrument : Roche Cobas Pure

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Lipase	U/L	32.70	29.53	0.61	0.33
@ Glucose	mg/dl	73.60	72.12	-0.21	-0.43
@ Urea	mg/dl	19.40	19.02	0.07	0.02
@ Creatinine	mg/dl	1.85	1.81	1.41	0.56
@ Bilirubin Total	mg/dl	3.75	3.77	-0.02	0.05
@ Bilirubin Direct	mg/dl	0.53	0.43	1.36	1.36
@ Total Protein	g/dL	5.70	5.49	0.20	0.12
@ Albumin	g/dL	3.47	3.46	0.18	0.22
@ Uric Acid	mg/dl	3.83	3.74	0.06	-0.04
@ Calcium	mg/dl	7.29	7.16	0.13	0.08
@ Phosphorous Inorganic	mg/dl	3.29	3.24	0.02	0.38
@ Cholesterol	mg/dl	135.00	133.86	0.44	0.44
@ Triglycerides	mg/dl	123.00	123.43	0.45	0.09
@ HDL	mg/dl	12.60	12.24	-1.18	-1.27
@ LDL	mg/dl	88.50	85.27	0.58	-0.04
@ ALP	U/L	48.00	46.86	0.48	0.26
@ GGT	U/L	23.20	21.57	0.55	0.31
@ Amylase	U/L	57.20	52.58	0.00	0.20
@ Chloride	mmol/L	78.40	76.82	-0.28	-0.25
@ Sodium	mmol/L	112.60	110.78	0.54	0.33
@ Potassium	mmol/L	3.80	3.74	0.54	0.41

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score ≥ 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	21
Not Evaluated Parameters	0
Evaluated Parameters	21
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070

Instrument : Roche Cobas Pure

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	73.00	73.20	-0.65	-0.43
@ Creatinine	mg/dl	0.73	0.71	0.29	0.56
@ Albumin	g/dL	3.98	4.00	0.36	0.22
@ Uric Acid	mg/dl	4.09	4.06	-0.13	-0.04
@ Calcium	mg/dl	8.01	7.98	0.00	0.08
* ALT	U/L	0.24	1.80	*	0.47
@ AST	U/L	17.40	17.63	-0.20	0.31
@ Chloride	mmol/L	97.00	93.83	-0.04	-0.25
@ Sodium	mmol/L	135.70	131.47	0.24	0.33
@ Potassium	mmol/L	5.12	4.87	0.37	0.41

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score ≥ 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	10
Not Evaluated Parameters	1
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A
 Kohinoor Mall, Kirod Road, Kurla (W),
 Mumbai - 400070