

Instrument : Horiba Yumizen H550

| Analyte | Standard Unit | Result Value | Mean | Z-Score | RMZ |
|--|---------------|--------------|---------|---------|-------|
| @ Erythrocyte (RBC) Count | mill/cu.mm | 1.07 | 1.02 | 0.20 | 0.08 |
| @ Total Leucocytes (WBC) count | /c.mm | 1830.00 | 1835.53 | -0.02 | -0.24 |
| @ Haemoglobin | g/dL | 2.70 | 2.68 | 0.00 | 0.05 |
| @ MCV (Mean Corpuscular Volume) | fL | 128.30 | 125.66 | 0.16 | 0.11 |
| @ MCH (Mean Corpuscular Hb) | pg | 25.40 | 26.65 | -0.25 | -0.11 |
| @ MCHC (Mean Corpuscular Hb Concentration) | g/dL | 19.80 | 20.97 | -0.32 | -0.17 |
| @ Platelet Count | 1000/ μ L | 97.00 | 87.19 | 1.11 | 1.28 |
| @ Hematocrit | % | 13.70 | 12.66 | 0.34 | 0.18 |
| @ RDW CV% | % | 23.90 | 24.07 | 0.05 | 0.61 |

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score \geq 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

| | |
|------------------------------|----------|
| Total Parameters | 9 |
| Not Evaluated Parameters | 0 |
| Evaluated Parameters | 9 |
| Outlier Parameters (X) | 0 |
| EQAS Score Haematology - CBC | 100.00 % |

DK Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070

Instrument : Horiba Yumizen H550

| Analyte | Standard Unit | Result Value | Mean | Z-Score | RMZ |
|--|---------------|--------------|---------|---------|-------|
| @ Erythrocyte (RBC) Count | mill/cu.mm | 1.48 | 1.45 | -0.05 | 0.08 |
| @ Total Leucocytes (WBC) count | /c.mm | 3240.00 | 3456.41 | -0.46 | -0.24 |
| @ Haemoglobin | g/dL | 4.10 | 4.05 | 0.10 | 0.05 |
| @ MCV (Mean Corpuscular Volume) | fL | 125.80 | 124.09 | 0.06 | 0.11 |
| @ MCH (Mean Corpuscular Hb) | pg | 27.50 | 27.88 | 0.04 | -0.11 |
| @ MCHC (Mean Corpuscular Hb Concentration) | g/dL | 21.90 | 22.18 | -0.02 | -0.17 |
| @ Platelet Count | 1000/ μ L | 174.00 | 151.44 | 1.45 | 1.28 |
| @ Hematocrit | % | 18.60 | 17.87 | 0.01 | 0.18 |
| @ RDW CV% | % | 25.60 | 22.46 | 1.17 | 0.61 |

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score \geq 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

| | |
|------------------------------|----------|
| Total Parameters | 9 |
| Not Evaluated Parameters | 0 |
| Evaluated Parameters | 9 |
| Outlier Parameters (X) | 0 |
| EQAS Score Haematology - CBC | 100.00 % |

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirod Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS
 REDCLIFFE LIFETECH PRIVATE LIMITED, GURGAON (1002)
 Outlier And Analyte Summary Report
 Result Details For Cycle No 230310 and Sample No 04

Report Date : 19/12/2023

Instrument : -

| Analyte | Standard Unit | Result Value | Accepted Value |
|---------------|---------------|--------------|----------------|
| @ Neutrophil. | - | 40-70% | 40-70% |
| # Lymphocyte. | - | >40% | |
| @ Eosinophil. | - | <6% | <6% |

| | | | | | | | | | | | |
|---|---|------------------|---|--------------------------|---|----------------------|---|------------------------|---|---|----------|
| Legend @ : Acceptable | <table border="1"> <tr> <td>Total Parameters</td> <td>3</td> </tr> <tr> <td>Not Evaluated Parameters</td> <td>1</td> </tr> <tr> <td>Evaluated Parameters</td> <td>2</td> </tr> <tr> <td>Outlier Parameters (X)</td> <td>0</td> </tr> <tr> <td>EQAS Score Haematology-PS Slide Count Qualitative</td> <td>100.00 %</td> </tr> </table> | Total Parameters | 3 | Not Evaluated Parameters | 1 | Evaluated Parameters | 2 | Outlier Parameters (X) | 0 | EQAS Score Haematology-PS Slide Count Qualitative | 100.00 % |
| Total Parameters | | 3 | | | | | | | | | |
| Not Evaluated Parameters | | 1 | | | | | | | | | |
| Evaluated Parameters | | 2 | | | | | | | | | |
| Outlier Parameters (X) | | 0 | | | | | | | | | |
| EQAS Score Haematology-PS Slide Count Qualitative | 100.00 % | | | | | | | | | | |
| Ⓢ : Acceptable | | | | | | | | | | | |
| X : Unacceptable | | | | | | | | | | | |
| # : Not Evaluated | | | | | | | | | | | |
| Ⓞ : Delayed Result Entry | | | | | | | | | | | |

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirod Road, Kurla (W),
 Mumbai - 400070