

Instrument : Lifotronic H-9

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ										
@ HbA1c	%	7.10	7.25	-0.40	-0.13										
Legend @ : Acceptable		<table border="1"> <tr> <td>Total Parameters</td> <td>1</td> </tr> <tr> <td>Not Evaluated Parameters</td> <td>0</td> </tr> <tr> <td>Evaluated Parameters</td> <td>1</td> </tr> <tr> <td>Outlier Parameters (X)</td> <td>0</td> </tr> <tr> <td>EQAS Score HbA1c</td> <td>100.00 %</td> </tr> </table>				Total Parameters	1	Not Evaluated Parameters	0	Evaluated Parameters	1	Outlier Parameters (X)	0	EQAS Score HbA1c	100.00 %
Total Parameters	1														
Not Evaluated Parameters	0														
Evaluated Parameters	1														
Outlier Parameters (X)	0														
EQAS Score HbA1c	100.00 %														
! : 2.0 < z score < 3.0 - Warning															
X : z score ≥3 - Unacceptable															
# : Not Evaluated															
⊙ : Delayed Result Entry															
* : Not considered for evaluation.															

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A
 Kohinoor Mall, Kirod Road, Kurla (W),
 Mumbai - 400070