

Erba Lachema s.r.o.



Brno, 10.11.2016

To,
Redcliffe Lifetech Private Limited.
M-16, ground floor, Sector 14, Gurugram.

Instrument Name: ERBA LAURA
Instrument Serial No: 1060-1-202209

We Erba Lachema SRO with registered office in Czech Republic, Karasek 2219/1D, 621 00 BMO, represented by its Production manager Veronika Pleskova here by confirm as Producer of LAURA instrument that this device is firmly calibrated by the manufacturer and does not require any regular calibration by the user, correct calibration of instrument is controlled by grey strips provided initially together with the instrument.

RNDr. Pavel Suchan Ph.D.

RNDr. Pavel Suchan Ph.D.

Global Product Manager, Urine Analysis

Erba Lachema s.r.o.

Karásek 1d, Brno 62100, Czech Republic



JKA

2024.03.06 08:46

DEKAPHAN LAURA
Seq.No: 0001
URINORM N
2024.03.06 08:46

.....
BLD NEG
BIL NFG
UBG NORM
KET NEG
GLU NEG
PRO NEG
PH 5
NIT NEG
LEU NEG
SG 1.030

QC TEST: OK

DEKAPHAN LAURA
Seq.No: 0002
URINORM P
2024.03.06 08:49

.....
*BLD 2+
*BIL 1+
*UBG 4+
*KET 2+
*GLU 3+
*PRO 3+
PH 9
*NIT POS
*LEU 3+
SG 1.000

Saha
06/03/24

QC TEST: OK