


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 240129A/N-118	Page	: 1 of 1
CALIBRATED ON.	: 30 / 01 / 2024	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 29 / 01 / 2025	C.P. No.	: ACS/CP/10
SRF No.& Dt.OF RECPT	: 240129A/N Dt. 29.01.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342724000000146F Dt. 02.02.2024	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. M-16, ground floor, old DLF Market, near SBI bank, Sector 14, Gurugram, Haryana		
ITEM CALIBRATED	: Centrifuge		
	Make : Remi	Range	: (0 to 6000) RPM
	L. Count : 1 RPM	ID No.	: RCL/CP/Cent-02
	Location : Clinical Pathology	S. No.	: ZIKN-43659
ENVIRONMENT CONDITION:	Temp. { 30°C ± 10°C } R.H. {30% to 80%}		

STANDARD INSTRUMENTS USED :

INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	G-230908-1-1	08/09/2024

RESULTS : Mechanical Discipline

* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)
1000	1001.3	- 1.3
2000	2002.5	- 2.5
4000	4004.5	- 4.5
6000	6006.2	- 6.2
(average of readings)		

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 4.68 rpm of reading at coverage Factor k=1.96.

*U.C. : Under Calibration.

1. Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
2. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
3. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
4. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
6. Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

