


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 240129A/N-107	Page	: 1 of 1	
<b>CALIBRATED ON.</b>	: 29 / 01 / 2024	Ref. Std.	: DKD R-5-7	
<b>CALIBRATION DUE.</b>	: 28 / 01 / 2025	C.P. No.	: ACS/CP/11	
<b>SRF No. &amp; Dt. OF RECPT</b>	: 240129A/N Dt. 29.01.2024	Phy. Inspection	: O.K.	
<b>ULR No. &amp; ISSUE DATE</b>	: CC342724000000135F Dt. 02.02.2024	Cal. At	: Site / Lab	
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. M-16, ground floor, old DLF Market, near SBI bank, Sector 14, Gurugram, Haryana			
<b>ITEM CALIBRATED</b>	: Digital Thermo Hygrometer			
	Make	: Digital Series	Range	: (0 to 50)°C, (10-99)%
	L. Count	: 0.1 °C, 1%RH	ID No.	: RCL/HEM/HGM-01
	Location	: Hematology	S. No.	: --
<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C + 4°C } R.H. {30% to 75%}			
<b>STANDARD INSTRUMENTS USED :</b>				
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>	<b>DUE DATE</b>
RH & Temperature Sensor with Indicator	ACS/RHTIS/01/20365872	DVG	CC352323000001002F	14/07/2024
<b>RESULTS</b>	: Thermal Discipline			
* U. C. INSTT. READS	STANDARD READING	ERROR	UNC(+)	
(°C)	15.3	14.90	0.40	2.1
	25.5	24.95	0.55	2.1
	40.7	39.96	0.74	2.1
(average of readings)				
(%)	29	29.93	-0.93	2.5
	60	59.93	0.07	2.5
	91	89.91	2.09	2.5
(average of readings)				

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is  $\pm 2.1^\circ\text{C}$ , 2.5% at coverage Factor  $k=1.96$ .

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confimty statement not provided since not required.

Prepared by

....End of Calibration Certificate....

