


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 240129A/N-105	<b>Page</b>	: 1 of 1	
<b>CALIBRATED ON.</b>	: 29 / 01 / 2024	<b>Ref. Std.</b>	: ISO:8655:6	
<b>CALIBRATION DUE.</b>	: 28 / 01 / 2025	<b>C.P. No.</b>	: ACS/CP/08	
<b>SRF No.&amp; Dt.OF RECPT</b>	: 240129A/N Dt. 29.01.2024	<b>Phy. Inspection</b>	: O.K.	
<b>ULR No. &amp; ISSUE DATE</b>	: CC342724000000133F Dt. 02.02.2024	<b>Cal. At</b>	: Site / Lab	
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. M-16, ground floor, old DLF Market, near SBI bank, Sector 14, Gurugram, Haryana			
<b>ITEM CALIBRATED</b>	: Micro Pipette			
	<b>Make</b>	: Thermo Scientific	<b>Range</b>	: (20 to 200) µl
	<b>L. Count</b>	: --	<b>ID No.</b>	: RCL/BIO/PIP-05
	<b>Location</b>	: Clinical Biochemistry	<b>S. No.</b>	: SH88518

**ENVIRONMENT CONDITION:** Temp. { 25°C ± 4°C } R.H. {30% to 75%}

### STANDARD INSTRUMENTS USED :

INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025

### RESULTS :

#### Mechanical Discipline

Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		20	19.77	0.23
200		100	99.64	0.36
		200	199.74	0.26

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 0.4 µl at coverage factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

