

## ADVANCE CALIBRATION SERVICES

S CC-3427



(ISO: 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.) Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com

## Calibration Certificate

CERTIFICATE NO.	: 240129A/N-112	Page	: 1 of 1
CALIBRATED ON.	: 29 / 01 / 2024	Ref. Std.	: DKD R-5-1
CALIBRATION DUE.	: 28 / 01 / 2025	C.P. No.	: ACS/CP/12
SRF No.& Dt.OF RECPT	: 240129A/N Dt. 29.01.2024	Phy. Inspectio	n : O.K.
ULR No. & ISSUE DATE	: CC342724000000140F Dt. 02.02.2024	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. M-16, ground floor, old DLF Market, Sector 14, Gurugram, Haryana	near SBI bank,	
ITEM CALIBRATED	: Multi Thermometer		
	Make : Thermocare	Range : (-50	to 300) <b>°C</b>
	L. Count : 0.1 °C	ID No. : RCL/	BIO/D.Temp-02

Location	. Oliffical Dio	CHCHISTIY C.IV.	•	
				7.75
ENVIRONMENT CONDITION: Te	emp { 25°C + 4°C}	R H {30% to 75%	3	

STANDARD INSTRUM	ENTS USED :		191 XVVI	
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Dig. Thermometer With RTD Sensor Digital Thermometer with Sensor	ACS/TIS/01/20C106914 ACS/TIS/02/ 201550503	BCL BCL	IALTH/00409/23 30070827	09/10/2024 20/07/2024

RESULTS	: ET Discip * U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)
	- 39.1	- 39.86	0.76
	0.2	0.01	0.19
	100.2	100.0	0.20
	299.2	299.9	- 0.70
	(average	e of readings)	

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 1.5°C at coverage Factor *k*=2.

\*U.C

## Under Calibration

- 1. Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above
- 2. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
   The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- 5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use
- Confirmity statement not provided since not required.



