


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 240129A/N-114	Page	: 1 of 1
CALIBRATED ON.	: 29 / 01 / 2024	Ref. Std.	: DKD R-5-1
CALIBRATION DUE.	: 28 / 01 / 2025	C.P. No.	: ACS/CP/12
SRF No.& Dt.OF RECPT	: 240129A/N Dt. 29.01.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000142F Dt. 02.02.2024	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. M-16, ground floor, old DLF Market, near SBI bank, Sector 14, Gurugram, Haryana		
ITEM CALIBRATED	: Multi Thermometer		
	Make : Thermocare	Range	: (-50 to 300) °C
	L. Count : 0.1 °C	ID No.	: RCL/SR/D.Temp-04
	Location : Sample Receiving Deptt.	S.No.	: --
ENVIRONMENT CONDITION:	Temp. { 25°C ± 4°C} R.H. {30% to 75%}		

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Dig. Thermometer With RTD Sensor	ACS/TIS/01/20C106914	BCL	IALTH/00409/23	09/10/2024
Digital Thermometer with Sensor	ACS/TIS/02/ 201550503	BCL	30070827	20/07/2024

RESULTS : ET Discipline			
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)	
- 39.5	- 39.85	0.35	
0.3	0.01	0.29	
100.5	100.2	0.30	
299.6	299.9	- 0.30	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.5°C at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

