

SVR

SVR BIOTECH

H.No : 9/107/3/2, GVR Heights, Nagaram Main Road,
Near Vijaya Hospital Secunderabad - 506 083.

GSTIN/UIN: 36AQQPS4180E1Z5

Mail : info@svrbiotech.com

CALIBRATION CERTIFICATE

INSTRUMENT NAME : URINE ANALYSER - YD 200
SERIAL NUMBER : 200190808005
INSTRUMENT IDENTIFICATION : EQP - 261
MANUFACTURER NAME : MR INTERNATIONAL HEALTHCARE
TECHNOLOGY CO LTD
IMPORTED AND MARKETED IN INDIA BY: SVR BIOTECH
STANDARD USED : SELF CALIBRATION MODE
STANDARD TRACEABILITY : IN HOUSE CERTIFICATE

CALIBRATION:

The instrument function well in all the systems check, self check and maintenance procedures. The equipment was calibration and the Red, Blue & Green (R,B & G) filters of all the parameters and top & bottom were found to be within the acceptable range of 333 ± 10 .

QC PERFORMANCE : ACCEPTABLE
SELF CALIBRATION : ACCEPTABLE
PERFORMANCE QUALIFICATION : PASSED
CALIBRATION DATE : 01/01/2024
CALIBRATION : PASSED
CALIBRATION VALID UP TO : 30/06/2024

K. RAJASHEKAR RAJU

K. Rajashekar Raju

TECHNICAL SERVICE ENGINEER



SVR For You.....



Scanned with OKEN Scanner