

CALIBRATION CERTIFICATE

Issued To: Swastik Diagnostic Center- Ultrasound Center in Dehradun
1st Floor Shop No.1,Sbd Tower, Main,
Sahastradhara Rd, Suman Nagar, Adhoiwala,
Dehradun, - 248013 Uttarakhand India

Certificate No:FL/C/M/05012024-C005
J.O.No:FL/C/M/05012024-005
ULR Code:CC273924100000084 F
S.R.F.:FL/SRF/20240105/B/AD/8
Issued Date:01/02/2024



Recommended Date for Next Calibration: 05/01/2025

Details of Instrument : Micro Pipette
Make/Model:Micro Lab
ID/S.No: YE20BA50070665
Range:5 to 50 μ L
Resolution:0.5 μ L
Condition :Working Satisfactory



Environmental Conditions : Temperature: (23 \pm 1) $^{\circ}$ C
Humidity: (50 \pm 10) % RH

Location of Calibration : At Lab (FARE Labs Pvt. Ltd.) Gurugram

Standard (s) used & Uncertainty : Micro Weighing Balance
Make/ Model/S.No : Mettler Toledo / XPE56/B617400630

Traceability of Standard (s) : NPL, India
(Cert. No.21021341/D1.01/C-016) Valid upto 23/06/2024

Methodology Used : As per Calibration Procedure(s) No. :FL/C/M/SOP-04

Date of receipt : 05/01/2024

Date(s) of Calibration : 05/01/2024

Authorized Signatory
C.S.Joshi, Director

NOTE: The laboratory accepts responsibility for content of this certificate. The results reported in this certificate pertain only to the device calibrated at the time of and under the stated conditions of measurement identified as above. this certificate shall not be reproduced except in full, without written approval of the laboratory. this certificate is intended of only guidance and not for legal purpose of for advertisement. Any complaint about this certificate should be communicated in writing within 7 days of its issue. Total liability of FARELABS Pvt. Ltd. will be limited to the invoiced amount only. This recipient of this report is requested to confirm the authenticity and genuineness through Email to Fare Labs.

In case of any complaint/feedback regarding the calibration of the device, please send an email at feedback@farelabs.com or call at +91 9599221227.

Digitally Signed

Issued To: Swastik Diagnostic Center- Ultrasound Center in Dehardun
 1st Floor Shop No.1,Sbd Tower, Main, Sahastradhara Rd, Suman Nagar, Adhoiwala, Dehradun, - 248013 Uttarakhand India

Certificate No:FL/C/M/05012024-C005
 J.O.No:FL/C/M/05012024-005
 ULR Code:CC273924100000084 F
 S.R.F.:FL/SRF/20240105/B/AD/8
 Issued Date:01/02/2024
 Recommended Date for Next Calibration: 05/01/2025



Volume | Discipline: Volume

| S. No. | Nominal Value | Measured volume (ml) at 27 °C | Error | Expanded Uncertainty |
|--------|---------------|--------------------------------|-----------|----------------------|
| | µL | µL | µL | µL |
| 1 | 5 | 4.987682 | -0.012318 | ±0.020 |
| 2 | 25 | 24.913551 | -0.086449 | ±0.090 |
| 3 | 50 | 49.947928 | -0.052072 | ±0.090 |

All measurement results are traceable to the International system of units (SI), ensured by comparison, directly or indirectly with National or International standards through unbroken chain of calibration. The reported expanded uncertainty of measurement is evaluated at approximately 95 % of confidence level with coverage factor $k = 2.11$.

DUC*: Device Under Calibration

Remarks:

- 1.Next date of calibration is recommended as per the request of the customer.
- 2.Calibration points are chosen as per request of the customer.



Authorized Signatory
 C.S.Joshi, Director

NOTE: The laboratory accepts responsibility for content of this certificate. The results reported in this certificate pertain only to the device calibrated at the time of and under the stated conditions of measurement identified as above. this certificate shall not be reproduced except in full, without written approval of the laboratory. this certificate is intended of only guidance and not for legal purpose of for advertisement. Any complaint about this certificate should be communicated in writing within 7 days of its issue. Total liability of FARELABS Pvt. Ltd. will be limited to the invoiced amount only. This recipient of this report is requested to confirm the authenticity and genuineness through Email to Fare Labs.
In case of any complaint/feedback regarding the calibration of the device, please send an email at feedback@farelabs.com or call at +91 9599221227.

----- End of Certificate -----

Digitally Signed