



CERTIFICATE OF IN-HOUSE CALIBRATION

Name of the Equipment: **Laboratory Centrifuge**

Equipment I. D.: **GTB/ICTC/Centrifuge**

Working Range: **1500-3000 rpm**

Manufacture: **Laboratory Centrifuge**

Calibrated on: **27/07/23**

Next Calibration due: **26/07/24**

Method: **Traceability Certificate**

Certificate No. **220915A/NN-105/ HDRL/ BRD MC, Gorakhpur**

Standard: **Tachometer ACS/DTM/01/211-0155V Godrej C-220912-6-111/09/2023**

Test condition: **Working**

Instrument	Reading					Mean
	1	2	3	4	5	
Tachometer	1560	1571	1573	1575	1590	1573.8
Centrifuge	1500	1500	1500	1500	1500	1500

Acceptable variation: +/- 10% with Standard Tachometer

Remarks: Instrument has passed the acceptance criteria (Yes/No)


Checked & Verified
Dr. Sanjay Kumar
Quality Manager
HIV Diagnostic Laboratory