



### CERTIFICATE OF IN-HOUSE CALIBRATION

Name of the Equipment: Refrigerator Thermometer

Equipment I. D: GTB/ICTC/Thermometer

Working Range: 2-8 deg C.

Manufacture: NA

Calibrated on: 27/07/23

Next Calibration due: 26/07/24

Method: Traceability Certificate

MLNMC/Micro/HDL/DTHM/01/SRL Certificate No. 230724B/NN Dt. 24.07.2023

Universal Calibrator ACS/UC/01/300561 Anshaankan AIPL/2022/0846 dt. 10/11/2023

Test condition: Working

Instrument	Reading					Mean
	1	2	3	4	5	
GTB/ICTC/Thermoemter	3.9	4.5	4.8	5.0	4.9	4.62
Thermometer (Ref. Thermometer)	3.4	4.0	4.5	4.9	4.6	4.28

Acceptable variation: +/- 1.0 deg Celsius

Remarks: Instrument has passed the acceptance criteria (Yes/No)

Checked & Verified by  
**Dr. Sanjay Kumar**  
Quality Manager  
HIV Diagnostic Laboratory