

Instrument : Erba ELite 580

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.13	2.04	0.64	0.13
@ Total Leucocytes (WBC) count	/c.mm	3820.00	3810.00	0.27	0.09
@ Haemoglobin	g/dL	6.10	5.95	0.74	0.06
@ MCV (Mean Corpuscular Volume)	fL	138.20	136.92	0.26	0.22
@ MCH (Mean Corpuscular Hb)	pg	28.60	29.35	-0.36	-0.26
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	20.70	21.45	-0.74	-0.43
@ Platelet Count	1000/ μ L	104.00	113.92	-0.76	-0.77
@ Hematocrit	%	29.50	27.81	0.67	0.28
@ RDW CV%	%	21.90	21.45	0.49	-0.12

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score \geq 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070

Instrument : Erba ELite 580

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.24	2.16	0.72	0.13
@ Total Leucocytes (WBC) count	/c.mm	8040.00	8336.36	-0.09	0.09
@ Haemoglobin	g/dL	7.00	6.83	0.31	0.06
@ MCV (Mean Corpuscular Volume)	fL	143.30	142.82	-0.02	0.22
@ MCH (Mean Corpuscular Hb)	pg	31.40	31.70	-0.08	-0.26
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	21.90	22.27	-0.13	-0.43
@ Platelet Count	1000/ µL	203.00	218.33	-0.93	-0.77
@ Hematocrit	%	32.10	30.58	0.54	0.28
@ RDW CV%	%	21.20	21.48	-0.51	-0.12

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %




Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
# Lymphocyte.	-	>40%	
@ Neutrophil.	-	40-70%	40-70%
@ Eosinophil.	-	<6%	<6%

Legend @ : Acceptable	<table border="1"> <tr> <td>Total Parameters</td> <td>3</td> </tr> <tr> <td>Not Evaluated Parameters</td> <td>1</td> </tr> <tr> <td>Evaluated Parameters</td> <td>2</td> </tr> <tr> <td>Outlier Parameters (X)</td> <td>0</td> </tr> <tr> <td>EQAS Score Haematology-PS Slide Count Qualitative</td> <td>100.00 %</td> </tr> </table>	Total Parameters	3	Not Evaluated Parameters	1	Evaluated Parameters	2	Outlier Parameters (X)	0	EQAS Score Haematology-PS Slide Count Qualitative	100.00 %
Total Parameters		3									
Not Evaluated Parameters		1									
Evaluated Parameters		2									
Outlier Parameters (X)		0									
EQAS Score Haematology-PS Slide Count Qualitative	100.00 %										
@ : Acceptable											
X : Unacceptable											
# : Not Evaluated											
⌚ : Delayed Result Entry											



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070