


Instrument : TransAsia Laura urine analyser

| Analyte | Standard Unit | Result Value | Accepted Value |
|------------------------|---------------|--------------|-------------------|
| @ pH | - | 6 | 5,5,5,6 |
| @ Specific Gravity | - | 1.025 | 1.020,1.025,1.030 |
| @ Glucose (urine) | - | 2+ | +2,+3,+4 |
| @ Protein (Urine) | - | 2+ | +1,+2,+3 |
| @ Ketone (urine) | - | Negative | Neg |
| @ Blood (urine) | - | 2+ | +1,+2,+3 |
| @ Bilirubin (urine) | - | Negative | Neg |
| @ Urobilinogen (urine) | - | Normal | Nor |
| @ Nitrite (urine) | - | Present | Pre |

Instrument : -

| Analyte | Standard Unit | Result Value | Accepted Value |
|--------------------------|---------------|--|----------------|
| @ Urine Pregnancy Test | - | Positive | Pos |
| Legend @ : Acceptable | | Total Parameters | 10 |
| @ : Acceptable | | Not Evaluated Parameters | 0 |
| X : Unacceptable | | Evaluated Parameters | 10 |
| # : Not Evaluated | | Outlier Parameters (X) | 0 |
| ⊙ : Delayed Result Entry | | EQAS Score Clinical Path - Urine Routine Chemistry | 100.00 % |



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070