



Date: 20' November 2023

To,  
Redcliffe Lifetech Pvt.Ltd

JAMMU (J&K)

Instrument Name: ERBA LAURA  
Instrument Serial no: 1020-1-202206

We Erba Lachema SRO with registered office in Czech Republic, Karasek 2219/1D, 621 00 BMO, represented by its Production manager Veronika Pleskova here by confirm as Producer of LAURA instrument that this device is firmly calibrated by the manufacturer and does not require any regular calibration by the user, correct calibration of instrument is controlled by grey strips provided initially together with the instrument.

12.9. 10/23

Erba Lachema s.r.o.  
Karasek 1d. 621 00 Brno  
DIČ CZ26918846 -10-

RNDr. Pavel Suchan Ph.D.  
Global Product Manager, Urine Analysis  
Erba Lachema s.r.o.  
Karasek 1d, Brno 62100, Czech Republic



Erba Lachema s.r.o. Karasek 1d. 621 00 Brno, Czech Republic  
Identification number: 269 18 846. Tax identification number: CZ26918846  
Incorporated in the Commercial Register maintained by the Register Court in Brno, Section C, insert 45436  
Tel: +420 517 077 111 fax: +420 517 077 077 e-mail: diagnostika@erbalachema.com www.erbalachema.com

DEKAPHAN LAURA

Seq No: 001

URINORM

2024.03.14

23:29

BIL NEG  
BLD NEG  
UBG NORM  
KET NEG  
GLU NEG  
PRO NEG  
pH 6  
NIT NEG  
LEU NEG  
SG 1.025

QC TEST: OK

DEKAPHAN LAURA

Seq.No: 0002

URINORM P

2024.03.14

23:30

*Handwritten signature and date: 14/3/24*

\*BIL 6 mg/dl  
\*BLD 250 Ery/u1  
\*UBG 12 mg/dl  
\*KET 52 mg/dl  
\*GLU 3+  
\*PRO 3+  
pH 9  
\*NIT POS  
\*LEU 500 Leu/u1  
SG 1.000

QC TEST: OK