


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 240201A/N-115	<b>Page</b>	: 1 of 1	
<b>CALIBRATED ON.</b>	: 03 / 02 / 2024	<b>Ref. Std.</b>	: IS:12508:1988	
<b>CALIBRATION DUE.</b>	: 02 / 02 / 2025	<b>C.P. No.</b>	: ACS/CP/10	
<b>SRF No.&amp; Dt.OF RECPT</b>	: 240201A/N Dt. 01.02.2024	<b>Phy. Inspection</b>	: O.K.	
<b>ULR No. &amp; ISSUE DATE</b>	: CC34272400000165F Dt. 05.02.2024	<b>Cal. At</b>	: Site / Lab	
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. Ward No.61, Indian Oil Petrol Pump, Adjoining Pizza Wings, Paloura Patta Bohri, Jammu-180002.			
<b>ITEM CALIBRATED</b>	: Centrifuge			
	<b>Make</b> : Remi	<b>Range</b>	: (0 to 6000) RPM	
	<b>L. Count</b> : 1 RPM	<b>ID No.</b>	: RCL/BIO/Cent-01	
	<b>Location</b> : Clinical Biochemistry	<b>S. No.</b>	: ZICN-14128	
<b>ENVIRONMENT CONDITION:</b>	Temp. { 30°C ± 10°C} R.H. {30% to 80%}			
<b>STANDARD INSTRUMENTS USED :</b>				
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>	<b>DUE DATE</b>
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	G-230908-1-1	08/09/2024
<b>RESULTS</b> : Mechanical Discipline				
<b>* U. C. INSTT. READS (RPM)</b>	<b>Standard Measured (RPM)</b>	<b>Error (RPM)</b>		
1000	1001.1	- 1.1		
2000	2002.2	- 2.2		
4000	4005.3	- 5.3		
6000	6006.1	- 6.1		
(average of readings)				

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is  $\pm 4.68$ rpm of reading at coverage Factor  $k=1.96$ .

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmy statement not provided since not required.

Prepared by

....End of Calibration Certificate....

