


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 240201A/N-116	Page	: 1 of 1
CALIBRATED ON.	: 03 / 02 / 2024	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 02 / 02 / 2025	C.P. No.	: ACS/CP/10
SRF No.& Dt.OF RECPT	: 240201A/N Dt. 01.02.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342724000000166F Dt. 05.02.2024	Cal. At	: Site / Lab

CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Ward No.61, Indian Oil Petrol Pump, Adjoining Pizza Wings, Paloura Patta Bohri, Jammu-180002.
-----------------------	--

ITEM CALIBRATED	: Centrifuge
Make	: Remi
L. Count	: 1 RPM
Location	: Clinical Pathology
Range	: (0 to 6000) RPM
ID No.	: RCL/CP/Cent-02
S. No.	: ZIEN-24019

ENVIRONMENT CONDITION:	Temp. {30°C ± 10°C} R.H. {30% to 80%}
-------------------------------	---------------------------------------

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	G-230908-1-1	08/09/2024

RESULTS :			Mechanical Discipline
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
1000	1001.3	- 1.3	
2000	2002.5	- 2.5	
4000	4004.7	- 4.7	
6000	6006.5	- 6.5	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 4.68rpm of reading at coverage Factor $k=1.96$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by 

....End of Calibration Certificate....

