


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 240201A/N-119	Page	: 1 of 1	
CALIBRATED ON.	: 03 / 02 / 2024	Ref. Std.	: CG-11	
CALIBRATION DUE.	: 02 / 02 / 2025	C.P. No.	: ACS/CP/07	
SRF No. & Dt. OF RECPT	: 240201A/N Dt. 01.02.2024	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342724000000169F Dt. 05.02.2024	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Ward No.61, Indian Oil Petrol Pump, Adjoining Pizza Wings, Paloura Patta Bohri, Jammu-180002.			
ITEM CALIBRATED	: Digital Temperature Indicator			
	Make	: RoHS	Range	: (-50 to 110) °C
	L. Count	: 0.1°C	ID No.	: RCL/BIO/D.Temp-01
	Location	: Clinical Biochemistry	S. No.	: -
ENVIRONMENT CONDITION:	Temp. { 30°C ± 10°C } R.H. {30% to 80%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Universal Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/041123/1-1	07/11/2024
RESULTS	: ET Discipline			
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)		
2.1	2	0.1		
4.2	4	0.2		
6.2	6	0.2		
8.4	8	0.4		
(average of readings)				

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=1.96.

*U.C.

: Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmy statement not provided since not required.

Prepared by

....End of Calibration Certificate....

