


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 240201A/N-101	Page	: 1 of 1
CALIBRATED ON.	: 01 / 02 / 2024	Ref. Std.	: ISO:8655:6
CALIBRATION DUE.	: 31 / 01 / 2025	C.P. No.	: ACS/CP/08
SRF No.& Dt.OF RECPT	: 240201A/N Dt. 01.02.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000151F Dt. 05.02.2024	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Ward No.61, Indian Oil Petrol Pump, Adjoining Pizza Wings, Paloura Patta Bohri, Jammu-180002.		
ITEM CALIBRATED	: Micro Pipette		
	Make : Thermo Scientific	Range	: (2 to 20) µl
	L. Count : --	ID No.	: RCL/HEM/PIP-01
	Deptt. : Hematology	S. No.	: TW08253
ENVIRONMENT CONDITION:	Temp. { 25°C ± 4°C } R.H. {30% to 75%}		

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
20	0.1	10	9.62	0.38
		20	19.81	0.19

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is $\pm 0.4 \mu\text{l}$ at coverage factor $k=2$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

...End of Calibration Certificate....

