


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 240201A/N-107	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 01 / 02 / 2024	<b>Ref. Std.</b>	: ISO:8655:6
<b>CALIBRATION DUE.</b>	: 31 / 01 / 2025	<b>C.P. No.</b>	: ACS/CP/08
<b>SRF No.&amp; Dt.OF RECPT</b>	: 240201A/N Dt. 01.02.2024	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC342724000000157F Dt. 05.02.2024	<b>Cal. At</b>	: Site / Lab
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. Ward No.61, Indian Oil Petrol Pump, Adjoining Pizza Wings, Paloura Patta Bohri, Jammu-180002.		
<b>ITEM CALIBRATED</b>	: Micro Pipette		
	<b>Make</b> : Erba	<b>Range</b>	: (10 to 100) µl
	<b>L. Count</b> : -	<b>ID No.</b>	: RCL/SER/PIP-07
	<b>Deptt.</b> : Serology	<b>S. No.</b>	: YE213AS0161050
<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C + 4°C } R.H. {30% to 75%}		

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
100	0.2	10	9.57	0.43
		50	49.67	0.33
		100	99.71	0.29

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is  $\pm 0.4 \mu\text{l}$  at coverage factor  $k=2$ .

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmy statement not provided since not required.

Prepared by

....End of Calibration Certificate....

