

Quality Control 1:					Exp: Mar 2024
Lot:- UQC2210A-L1					
S.No.	Na+	K+	iCa	Cl ⁻	
1	121.4	2.10	1.92	79.8	
2	121.3	2.08	1.89	80.5	
3	121.6	2.07	1.85	81.0	
Mean	121.4	2.08	1.88	80.4	
Ranges	113-128 mmol/L	1.5-2.5mmol/L	1.25-2.25mmol/L	76-90mmol/L	
Status	O.K	O.K	O.K	O.K	
Quality Control 2:					Exp: Mar 2024
Lot:- UQC2210A-L2					
S.No.	Na+	K+	iCa	Cl ⁻	
1	140.0	3.99	1.22	99.5	
2	139.5	4.01	1.19	100.1	
3	139.5	3.98	1.20	99.8	
Mean	139.6	3.99	1.20	99.8	
Ranges	132-153mmol/L	3.5-4.7mmol/L	0.62-1.62mmol/L	95-111mmol/L	
Status	O.K	O.K	O.K	O.K	
Quality Control 3:					Exp: Mar 2024
Lot:- UQC210A-L3					
S.No.	Na+	K+	iCa	Cl ⁻	
1	160.50	5.94	0.70	120.1	
2	161.05	5.90	0.71	120.6	
3	161.50	5.93	0.72	121.2	
Mean	161.01	5.92	0.71	120.6	
Ranges	152-164mmol/L	5.4-6.8mmol/L	0.75-1.15mmol/L	116-128mmol/L	
Status	O.K	O.K	O.K	O.K	



Signature of Engineer

DATE: 14/01/2024

Signature of Doctor

DATE: