



FACTUAL CALIBRATION SERVICES (OPC) PVT. LTD.

AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY

A House of Complete Solution

Contact Details:- +91-9205523341, +91-9868784023

Office No. 130, 1st Floor Cloud-9, Sec.-1, Vaishali, Ghaziabad, Uttar Pradesh-201010



Calibration Certificate

Format No. : FCS/QF/46

ISSUE DATE 01/07/2023

Page No. : 01 of 01

SERVICE REQUEST & CALIBRATION REPORT DETAILS (As per agreed with the Customer)

Calibration Request Dated	01.03.2024	Date of Calibration :	02.03.2024
Calibration Request / Job No.	227 / 227-005	Next Date of Calibration	02.03.2025
Certificate No.	FCS/24/ 227-005	Certificate Issue Date	03.03.2024

CUSTOMER ADDRESS	ENVIRONMENTAL & INSTRUMENT CONDITIONS
M/s ARUNOSHA LABS PVT. LTD. 203/4, Local Shopping Center Gujranwala Town, Part-I New Delhi 110009 INDIA	Temperature (25 ± 15) °C Relative Humidity (50 ± 20) %RH Zero Error Null Discipline Medical Instrument Condition In Working (Good)

CALIBRATION CERTIFICATE OF	DESCRIPTION OF(EQP) UNIT UNDER CALIBRATION
Micropipette	Make / Model DRAGON LAB Range / Size 10 to 100 µl Least Count 0.5 µl Serial No. NA Asset Code / I.D No. MP-03 Location LAB Accuracy/Acceptance Not Specified
Procedure Followed FCS/MD/120	
Reference Standard SO 8655-6	
Calibration Performed At In Lab	

DETAILS OF REFERENCE STANDARDS AND MAJOR EQUIPMENTS USED FOR CALIBRATION

Sr. No	Instrument Details	Asst Code	Certificate No.	Calibrated By	Valid upto
1	Digital Weighing Machine	FCS/MD/PWB-01	NKERC/23/1001-012	NKERC	01.08.2024

The Standards used for calibration are traceable to National Standards.

CALIBRATION RESULT

PARAMETER	DUC VALUE (µl)	STANDARD VALUE (gm)	STANDARD VALUE Converted (µl)	ERROR (µl)	Uncertainty ± (µl)
Volume	20	0.00199	19.90	0.10	4.0
	50	0.04954	49.70	0.30	4.0
	60	0.05987	59.64	0.36	4.0
	70	0.06999	69.58	0.42	4.0
	100	0.09985	99.40	0.60	4.0

*****The End of Report*****

CALIBRATED BY (Calibration Engineer)	APPROVED BY (Authorized Signatory)
--	--

Note :-	<ol style="list-style-type: none"> 1. UUC =* Unit Under Calibration* 2. Uncertainty of measurement at approximately 95% Confidence Level with coverage factor k=2, ±20 µm 3. The Calibration Result reported are valid at the time of Calibration and under the stated conditions of Measurement. 4. The Calibration report is not valid without sign of CEO, OPC, Factual Calibration Services (OPC) PVT. LTD.. 5. This Calibration Certificate only to the particular Equipment / Instrument submitted for Calibration. 6. The Calibration Certificate Should not be regenerated without the permission of Higher Authority of Factual Calibration Services (OPC) PVT. LTD.. 7. Calibration Certificate issue for weight & Measure parameters i.e. Mass, Balance, Volumetric equipment, for scientific Purpose only.
----------------	---

A CALIBRATION LABORATORY

www.factualcalibrationpvtltd.com

factualcalibration23@gmail.com



सूक्ष्म, लघु एवं मध्यम उद्यम
MICRO, SMALL & MEDIUM ENTERPRISES

OUR BRANCHES: GHAZIABAD KATHMANDU LUCKNOW REWARI NOIDA PUNJAB