



UNIQUE CALIBRATION & CONSULTANT SERVICES

AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY



CC-3287

CALIBRATION CERTIFICATE

Format No.: UCCS-QF-7.8.1	Certificate No.: UCCSCC20230213-002	ULR No.: CC328723000002498F			
Group: Volume	Discipline: Mechanical				
Customer's Name & Address M/s. Dr. Bharti Path Lab Centre 1, Opp. PNB Bank, Mohna Road, Ballabhgarh	Page No.: 1 of 1	Calibration Performed: At Lab			
	Date of Receipt: 13-02-2024				
	Service Request No.: UCCSSRN20230213-001				
	Certificate Issue Date: 15-02-2024				
	Date of Calibration: 14-02-2024				
Next Due Date (Suggested by Customer): 13-02-2025					
Equipment Description (UUC) & Identification					
Name of the Instrument: Micro Pipette	Make: Erba				
	Model: -----				
Range: 100 µl to 1000 µl	Sr. No.: -----				
Least Count: 5 µl	I.D. No.: MP-02				
Condition of Instrument: Ok	Location: -----				
Master Equipment Used					
Sr. No.	Instrument Name	Range & I.D. No.	Calibration Agency	Calibration Certificate No.	Valid Up to
1	Weight Set (E1 Class)	1 mg to 200 g & UCCS/CAL/048	Weightronics	WMCL/E/2022-11/2116	07-11-2025
2	Digital Weighing Balance	0 to 220 g & UCCS/CAL/005	Parashar Micro Measurement Pvt. Ltd.	PMM/101122/14-1	07-11-2025
Environmental Conditions			Calibration Procedure	Calibration Reference Standard	
Temperature: 20 ± 2 °C, Humidity: 50 ± 10 %RH			UCCS 03 SOP 018	ISO 8655	
Atmospheric Pressure : 14.2 psi					

CALIBRATION RESULTS

Sr. No.	Nominal Capacity (µl) at 20 °C	Actual Volume (µl) at 20 °C	Error (µl)	Expanded Uncertainty (±) µl
1	100	101.61	-1.61	4.8
2	500	502.99	-2.99	4.8
3	1000	1004.38	-4.38	4.8

End of Certificate

Hemraj
Calibrated By

Hemraj

Calibration Engineer

Reviewed/Approved By
L.M. Mahato
Chief Executive Officer (CEO)

Note: - 1. The calibration results reported in this certificate are valid at the time of and stated condition of measurement. 2. This report is not to be reproducing wholly or in part and cannot be used as evidence in the Court of Law and should not be used in any advertising media without our special permission in writing. 3. Total liability of our Organization is limited to the invoiced amount. 4. Calibration Certificate without signature are not valid. 5. All our certificates are metrological traceable. 6. The results relate only to the item calibrated. 7. UUC: Unit Under Calibration. 8. Expanded Uncertainty at 95 % confidence level with coverage factor K=2. 9. Scan this QR code to download our NABL Accreditation Scope & Certificate.	
--	--

Working : W.Z.1390/Z-2, 1st Floor, Nangal Raya Extn., Pankha Road, New Delhi-110046

Regd. Office : R.Z.56A, Indra Park, Gali No.25, Palam Colony, New Delhi-110045
M.: 9958445875, 9313768426, 9999761885 E-Mail : uniquecalibrationservices@gmail.com