



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - FEBRUARY 2024

OPPO F17



PC-1024

Lab Name	Govt. PHC, Muthalipalayam	Lab No	12887
Constituent Group	Chemistry III	Date of Result Entered :	20/02/2024
PT item	Lyophilized human serum based	Date of Report Published :	05/03/2024

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1924	97.33	5.78	5.62	91.8 mg/dL	-0.98	0.26
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1884	2.18	6.74	0.15	2.3 mg/dL	0.82	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1800	113.08	5.21	5.90	109 mg/dL	-0.69	0.28

u\* - Method of Uncertainty

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  \geq 3.0$	Unacceptable (action Signal)

**LAB ADDRESS :**

Govt. PHC, Muthalipalayam  
GOVT PRIMARY HEALTH CENTER  
Muthalipalayam  
TAMILNADU641606

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components