



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JANUARY 2024



PC-1024

OPPO F17

Lab Name: Govt. PHC, Muthalpalayam
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based

Lab No: 12887
 Date of Result Entered: 23/01/2024
 Date of Report Published: 06/02/2024

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1882	145.08	6.30	9.14	150 mg/dL	0.54	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1872	1.64	7.16	0.12	1.6 mg/dL	-0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1799	107.07	5.63	6.03	109 mg/dL	0.32	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. PHC, Muthalpalayam
 GOVT PRIMARY HEALTH CENTER
 Muthalpalayam
 TAMILNADU641606

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Confidential Data in CMC EQAS reports is confidential