

CHRISTIAN MEDICAL COLLEGE

OPPO F17



DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2023

PC-1024

Lab Name Govt. PHC, Muthalpalayam

Lab No

12887

Constituent Group Chemistry III

Date of Result Entered :

21/12/2023

PT item Lyophilized human serum based

Date of Report Published :

02/01/2024

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1885	99.99	6.05	6.05	96 mg/dL	-0.66	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1844	1.74	6.08	0.11	1.8 mg/dL	0.57	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	107.76	4.90	5.28	101 mg/dL	-1.28	0.25

u* - Method of Uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

Govt. PHC, Muthalpalayam
GOVT PRIMARY HEALTH CENTER
Muthalpalayam
TAMILNADU641606

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

A. M.
மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலையம்
முதலியானம்,
தீருப்பூர் - 641 606.