



AACALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE



CC 2646

Certificate No	AACPL / 07053F	Field	Mech	Group	Volume	Page No.	1 of 1
SRF No & Date	230926-4 DT - 26 / 09 / 2023			ULR No.	CC2646230000 07053F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd Mathura Daas Civil Hospital (DH). Main Bazar, Moga (PB)			Calibration Date	26-09-2023		
Address				Recommended Date for Next Callb.	25-09-2024		
				Certificate Issue Dt.	29-09-2023		
				UUC Condition	Good		

Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	10 to 100 µl	1 µl	Dragon Lab	KDDHPBCCMP014 Bio Chemistry

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6	Volume Measurement	(50 ± 10)%RH	(23 ± 2)°C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No. CP - 03

Calibration Result

Sl. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor k = 2 is ± (in µl)
1	10	9.99856	0.00144	0.25
2	50	49.99637	0.00363	0.25
3	100	99.99421	0.00579	0.25

- Note**
- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
 - 2 This certificate cannot be reproduced except in full without our prior permission in writing
 - 3 This certificate refers only to the particular item(s) submitted for calibration
 - 4 UUC - Unit under calibration

----- End of Certificate -----

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Del. ch



Approved by