



ADL/BO/MS/C/0379/23-24

CERTIFICATE OF CALIBRATION

Instrument: MISPA CXL PRO PLUS Fully Automated Biochemistry Analyser
Customer Name: Amruth Diagnostic Centre, Bangalore
Date of Installation: 07-07-2022
S/N: 21T240PCSS0066K

Alignment Check

Motion Check	Checked	Remark
Sample Probe	<input checked="" type="checkbox"/>	OK
Reagent Probe	<input checked="" type="checkbox"/>	OK
Mixing Bar	<input checked="" type="checkbox"/>	OK
Reaction Disk	<input checked="" type="checkbox"/>	OK
Reagent Disk	<input checked="" type="checkbox"/>	OK
Sample Disk	<input checked="" type="checkbox"/>	OK
Cuvette Check	<input checked="" type="checkbox"/>	OK
Lamp	<input checked="" type="checkbox"/>	OK

System	Checked	Remark
Main unit	<input checked="" type="checkbox"/>	OK
Temp. unit	<input checked="" type="checkbox"/>	OK
Mixing unit	<input checked="" type="checkbox"/>	OK
Reaction unit	<input checked="" type="checkbox"/>	OK
Reagent unit	<input checked="" type="checkbox"/>	OK
Sample unit	<input checked="" type="checkbox"/>	OK

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

MUMBAI (REGISTERED OFFICE)
 401 & 402, 4th Floor, Jaisirgh Business Centre, 119, Sahar Road, Partimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
 Tel: +91 484 286 7000 | Email: agappe@agappe.in

DELHI OFFICE
 DSM 540, 5th Floor, DLF Tower, Shivaji Marg, New Delhi - 110 015, India.
 Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
 5-6, 11th Floor, Red Cross Bhavan, No.26 Race Course Road, Bangalore - 560 001. Tel: +91 80 2228 8288
 Email: bangaloreoffice@agappe.in

KOLKATA OFFICE
 406, Merlin Matrix, Plot No-10, Block-DN, Sector V, Salt Lake City, Kolkata - 700 091.
 Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

EQUIPMENT PLANT
 X/588-CB, Block No. 32, KINFRA Small Industrial Park, Nellikud, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

YOUR BEST PARTNER IN DIAGNOSTICS



OPTICAL SECTION

Filter	Light Intensity	
340	10329	OK
380	10382	OK
405	10300	OK
450	10317	OK
480	10349	OK
505	10373	OK
546	10395	OK
570	10361	OK
600	10234	OK
660	10391	OK
700	10376	OK
800	10398	OK

Power Supply : 12V 24V

The equipment is calibrated for the different parameters by using Multi Calibrator

Lot.No: 33020264 Exp: 31-01-2025

This is to certify that the above mentioned product has been calibrated according to the standard operating procedure given by the Agappe Diagnostics Ltd.

Date of Calibration : 01/Jan/2024

Next Due : 30/June/2024

Agappe Diagnostics Ltd.,



Authorized Signatory
Name : Murugan Saravanan

Designation : Assistant Manager - Customer Service
AGAPPE DIAGNOSTICS LTD.

ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC15413

CORPORATE OFFICE / REAGENT PLANT
Pape Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

KOLKATA OFFICE
406, Merin Matrix, Plot No-10, Block-DIV, Sector-V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

BANGALORE OFFICE
5-6, 1st Floor, Bed Cross Bhawan, No.26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 6288
Email: bangaloreoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
402, 4th Floor, Jaishigh Business Centre, 119,
Mar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in