

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@Total Leucocytes (WBC) count	/c.mm	1660	2047.65	-1.16	-1.05
@ Erythrocyte (RBC) Count	mill/cu.mm	1.97	2.03	-0.59	-0.18
@ Haemoglobin	g/dL	5.50	5.47	0.05	0.09
@ MCV (Mean Corpuscular Volume)	fL	118.80	117.12	0.14	0.16
@ MCH (Mean Corpuscular Hb)	pg	27.90	26.96	0.66	0.30
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	23.50	23.00	0.24	0.07
@ Platelet Count	1000/ μ L	123.00	130.07	-0.01	0.25
@ Hematocrit	%	23.40	23.76	-0.25	-0.01
@ RDW CV%	%	18.30	19.25	-0.47	-0.09

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score \geq 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam

PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kiroli Road, Kurla (W),

Mumbai - 400070

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@Total Leucocytes (WBC) count	/c.mm	4910	5806.22	-0.94	-1.05
@ Erythrocyte (RBC) Count	mill/cu.mm	2.16	2.14	-0.03	-0.18
@ Haemoglobin	g/dL	6.30	6.22	0.20	0.09
@ MCV (Mean Corpuscular Volume)	fL	126.90	122.79	0.36	0.16
@ MCH (Mean Corpuscular Hb)	pg	29.30	29.02	0.27	0.30
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	23.10	23.71	-0.10	0.07
@ Platelet Count	1000/ μ L	221.00	227.36	-0.18	0.25
@ Hematocrit	%	27.40	26.07	0.19	-0.01
@ RDW CV%	%	16.70	18.24	-0.88	-0.09

Legend @ : Acceptable

| : 2.0 < z score < 3.0 - Warning

X : z score \geq 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam

PT coordinator & Technical-Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirod Road, Kurla (W),

Mumbai - 400070

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Neutrophil.	-	>70%	>70%
@ Lymphocyte.	-	<20%	<20%
@ Eosinophil.	-	<6%	<6%

Legend @ : Acceptable	<table border="1"> <tr> <td>Total Parameters</td> <td>3</td> </tr> <tr> <td>Not Evaluated Parameters</td> <td>0</td> </tr> <tr> <td>Evaluated Parameters</td> <td>3</td> </tr> <tr> <td>Outlier Parameters (X)</td> <td>0</td> </tr> <tr> <td>EQAS Score Haematology-PS Slide Count Qualitative</td> <td>100.00 %</td> </tr> </table>	Total Parameters	3	Not Evaluated Parameters	0	Evaluated Parameters	3	Outlier Parameters (X)	0	EQAS Score Haematology-PS Slide Count Qualitative	100.00 %
Total Parameters		3									
Not Evaluated Parameters		0									
Evaluated Parameters		3									
Outlier Parameters (X)		0									
EQAS Score Haematology-PS Slide Count Qualitative	100.00 %										
@ : Acceptable											
X : Unacceptable											
# : Not Evaluated											
Ⓢ : Delayed Result Entry											

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A
 Kohinoor Mall, Kirod Road, Kurla (W),
 Mumbai - 400070